

EC DECLARATION OF CONFORMITY

6.18 Declaration of Conformity

The EC declaration of conformity will be signed before the product is placed on the market and may be referred to in the Quality Document Management system as follows:

Document number	Document Title	Content
VENAGRADO-21	Declaration of conformity	The EC declaration of conformity is the written statement and the single declaration drawn up by the manufacturer to demonstrate the fulfilment of the EU requirements relating to a product bearing the CE marking he has manufactured. The declaration is in respect of all Community acts applicable to the product containing all information required for the identification of Community harmonisation legislation to which the declaration relates.

The following standards and regulations were applied in the documents related to this section:

Directive / Guideline	Title
Council Directive 93/42/EEC	14 June 1993 concerning medical devices
MEDDEV 2.4/1 rev 9	Classification of Medical Devices

EC DECLARATION OF CONFORMITY

Legal Manufacturer:	Speciality Fibres and Materials Ltd Galaxy House – 31 Herald Way, Binley Industrial Estate, CV3 2RQ, Coventry, United Kingdom
SRN:	GB-MF-000004153
Intended Use:	DURAFIBER Ag PRO antimicrobial wound dressings can be used for the management of moderately to heavily exuding wounds.
General Product Name/s:	DURAFIBER Ag PRO
Variants:	As per Annex I – Product Listing
Applicable Standards:	As Per Annex II
Classification:	Class III - Rule 13
GMDN Code/Term:	47045
Basic UDI-DI:	N/A
Conformity Assessment Procedure:	EC Declaration of Conformity in accordance with Annex II of the Medical Device Directive certificate number 1984-MDD-21-752 SFM Annex II.3
Notified Body:	Kiwa Belgelendirme Hizmetleri A.Ş. (CE1984) Tepeören Mevkii Ankara Asfaltı Maret Arkası ITOSB 9. Cadde No: 15 Tuzla, Istanbul Turkey
Authorised Representative:	Advena Ltd Tower Business Centre, 2nd Floor, Tower Street, Swatar, BKR 4013 Malta
EC Certificate Number:	1984-MDD-21-769
Start of CE-Marking:	5 th July 2022

This declaration of conformity is issued under the sole responsibility of Speciality Fibres and Materials Ltd.

Signature: 

Date and Place: 9th April 2024, Coventry

Name: Nyerngoor Korda Hewitt

Position: Director of RAQ

EC DECLARATION OF CONFORMITY

ANNEX I – Product Listing

Product Models (Commercial Names (Brands))	Product Reference	Smith and Nephew Re-OrderCode	Product Variants (Commercial Names/ Dimensions)
DURAFIBER Ag PRO	FV000971	66807678	DURAFIBER Ag PRO 10x10cm
DURAFIBER Ag PRO	FV000991	66807679	DURAFIBER Ag PRO 15x15cm
DURAFIBER Ag PRO	FV000995	66807677	DURAFIBER Ag PRO 5x5cm
DURAFIBER Ag PRO	FV000997	66807680	DURAFIBER Ag PRO 20x30cm
DURAFIBER Ag PRO	FV000999	66807681	DURAFIBER Ag PRO 2x45cm
DURAFIBER Ag PRO	FV001000	66807683	DURAFIBER Ag PRO 4x10cm
DURAFIBER Ag PRO	FV001001	66807684	DURAFIBER Ag PRO 4x20cm
DURAFIBER Ag PRO	FV001002	66807685	DURAFIBER Ag PRO 4x30cm

EC DECLARATION OF CONFORMITY

ANNEX II – List of Standards

This present declaration is also in conformity with the following European standards and Common Specifications:

Number	Standards Title	Version
EN ISO 11137-1	Sterilization of health care products - Radiation - Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices	2006, Amd 1:2013 2015, Amd 2:2019
EN ISO 11137-2	Sterilization of health care products - Radiation - Part 2: Establishing the sterilization dose	2013, 2015
ISO 11137-3	Sterilization of health care products - Radiation - Part 3: Guidance on dosimetric aspects	2017
EN ISO 10993-1	Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process	2018
EN ISO 10993-3	Biological evaluation of medical devices - Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity	2014
EN ISO 10993-5	Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity	2009
ISO 10993-6	Biological evaluation of medical devices - Part 6: Tests for local effects after implantation	2016
ISO 10993-10	Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization	2021
EN ISO 10993-11	Biological evaluation of medical devices - Part 11: Tests for systemic toxicity	2018
EN ISO 10993-12	Biological evaluation of medical devices - Part 12: Sample preparation and reference materials	2021
EN ISO 10993-17	Biological evaluation of medical devices - Part 17: Establishment of allowable limits for leachable substances	2023
EN ISO 10993-18	Biological evaluation of medical devices - Part 18: Chemical characterization of materials	2020
EN ISO 11607-1	Packaging for terminally sterilized medical devices - Part 1: Requirements for materials, sterile barrier systems and packaging systems	2020
EN ISO 11607-2	Packaging for terminally sterilized medical devices - Part 2: Validation requirements for forming, sealing and assembly processes	2020
EN ISO 11737-1	Sterilization of medical devices - Microbiological methods - Part 1: Determination of a population of microorganisms on products	2018
EN ISO 11737-2	Sterilization of medical devices - Microbiological methods - Part 2: Tests of sterility performed in the definition, validation and maintenance of a sterilization process	2020
EN ISO 13485	Medical devices - Quality management systems - Requirements for regulatory purposes	2016 + AC:2018
EN ISO 14971	Medical devices - Application of risk management to medical devices	2019

EC DECLARATION OF CONFORMITY

EN ISO 15223-1	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements	2021
BS EN 62366-1	Medical devices — Part 1: Application of usability engineering to medical devices	2015+ A1:2020
EN ISO 13726-1	Test methods for primary wound dressings - Part 1: Aspects of absorbency	2002 + AC:2003
ISO 14644-1	Cleanrooms and associated controlled environments. Classification of air cleanliness	2015
ISO 14644-2	Cleanrooms and associated controlled environments. Specifications for testing and monitoring to prove continued compliance with ISO 14644-1	2015
EN 1041	Information supplied by the manufacturer of medical devices	2008
EN 556-1	Sterilization of medical devices. Requirements for medical devices to be designated "STERILE" Requirements for terminally sterilized medical devices	2001
ASTM F1980-16	Standard Practice for Performance Testing of Shipping Containers and Systems	2016
ASTM D 4169-16	Standard Guide for Accelerated Aging of Sterile Barrier Systems for Medical Devices	2016