	Effective Date:	24-Jul-2020	CR:	CR-040524
	Status:	Effective	Version:	2.0, CURRENT
	Group:	Change Control	Sites:	CC-Deeside-R&D
	Artifact Name:	WI - Work Instruction	Name:	WI-0506
	Title:	Declaration of Conformity under EU Medical Device Regulation 2017/745		

EU Declaration of Conformity


Esenta™ Sting Free Skin Barrier Products


Technical Documentation MDR OSTTF 017


This Declaration of Conformity is issued under the sole responsibility of ConvaTec Limited.

We hereby declare that the mentioned product/ attached list of product comply with the applicable general safety and performance requirements and provisions of EU Medical Device Regulation (EU) 2017/745.

Legal Manufacturer:	ConvaTec Limited First Avenue, Deeside Industrial Park, Deeside, Flintshire CH52NU United Kingdom 01244 584000
SRN:	GB-MF-000001770
Authorized Representative:	Unomedical A/S Aaholmvej 1-3, Osted, 4320 Lejre Denmark
Product Name:	Refer to List of Product Codes attached
GMDN Code and Title:	58978- Synthetic-polymer liquid barrier dressing, non-sterile
CND nomenclature:	A108004- protective films / sprays for peristomal skin
Basic UDI-DI:	Refer to List of Product Codes attached
Identification of the device(s) concerned:	Refer to List of Product Codes attached
Catalogue Number:	Refer to List of Product Codes attached
Intended purpose:	Protection of skin from the damage associated with body waste, enzymes and adhesives.
Risk Classification:	Esenta™ Sting Free Skin Barrier Spray 50ml: Class I as per Rule 13 in Annex VIII Esenta™ Sting Free Skin Barrier Pump Spray 28ml and Wipes (25 and 30 pack): Class I as per Rule 1 in Annex VIII


	Effective Date:	24-Jul-2020	CR:	CR-040524
	Status:	Effective	Version:	2.0, CURRENT
	Group:	Change Control	Sites:	CC-Deeside-R&D
	Artifact Name:	WI - Work Instruction	Name:	WI-0506
	Title:	Declaration of Conformity under EU Medical Device Regulation 2017/745		

Conformity Assessment Route:	Annex II and III of EU Medical Device Regulation (EU) 2017/745
Relevant Harmonized Standards:	N/A, at time of approval there are no Harmonized Standards to the Medical Device Regulation 2017/745
References to any CS:	Not Applicable
Identification of Notified Body:	Not Applicable - Class I non-sterile device
Identification of the Certificate(s):	EC Quality Management System Certificate No. MD 670405, issued by BSI.
Identification of the person authorized to sign on behalf of Legal Manufacturer:	<p>Name: Gary Barrett</p> <p>Signature: </p> <p>Vice President, Regulatory Affairs</p> <p>Place of Issue: Deeside., Flintshire</p> <p>Date: august 18 2021</p>

	Effective Date:	24-Jul-2020	CR:	CR-040524
	Status:	Effective	Version:	2.0, CURRENT
	Group:	Change Control	Sites:	CC-Deeside-R&D
	Artifact Name:	WI - Work Instruction	Name:	WI-0506
	Title:	Declaration of Conformity under EU Medical Device Regulation 2017/745		

List of Product Codes

Product ICC	Product SAP	Trade Name	Basic UDI-DI
423288	1729205	Esenta™ Sting Free Skin Barrier Spray 50ml	768455OST0026FD
423286	1729201	Esenta™ Sting Free Skin Barrier Pump Spray 28ml	768455OST0027FF
423392	1733700	Esenta™ Sting Free Skin Barrier Wipes (1x25pack)	768455OST0025FB
423282	1729195	Esenta™ Sting Free Skin Barrier Wipes (1x30pack)	768455OST0025FB

	Effective Date:	24-Jul-2020	CR:	CR-040524
	Status:	Effective	Version:	2.0, CURRENT
	Group:	Change Control	Sites:	CC-Deeside-R&D
	Artifact Name:	WI - Work Instruction	Name:	WI-0506
	Title:	Declaration of Conformity under EU Medical Device Regulation 2017/745		

Revision History

Rev no	Date	Comment	Author
1.0	18/08/2021	Initial Release of Technical Documentation under the Medical Device Regulation (EU) 2017/745	Sana Ashraf