

**Konformitätserklärung
Declaration of Conformity**

Wir

We

B. Braun Melsungen AG
Carl-Braun-Str. 1
34212 Melsungen
Deutschland/Germany
SRN DE-MF-000000201erklären in eigener Verantwortung,
dass das Produkt / die Produkte**Perifix® LOR**
Perifix® LOR NRFit®Spritze für die Loss-of-Resistance (LOR)-
Methode in der Regionalanästhesie
(Artikelnummern und Basic UDI-DI siehe Anlage I)mit den Anforderungen der Medizinprodukte
Verordnung (EU) 2017/745 übereinstimmen**Konformitätsbewertungsverfahren**
nach Anhang IX**Klassifizierung**
gemäß Anhang VIII der oben genannten
Verordnung
Klasse I steril**Benannte Stelle**
TÜV SÜD Product Service GmbH
Kennnummer 0123**Gültig bis 2026-12-14**
gemäß gültigem EU-Zertifikat
(Nr. G11 012974 0626)hereby declare in our sole responsibility
that the product/s**Perifix® LOR**
Perifix® LOR NRFit®Loss of Resistance (LOR) device for use in
Regional Anesthesia
(article numbers and Basic UDI-DI see attachment I)are in conformity with the requirements of the
Medical Device Regulation (EU) 2017/745**Conformity Assessment Procedure**
according to annex IX**Classification**
according to annex VIII of the Regulation named
above
Class I sterile**Notified Body**
TÜV SÜD Product Service GmbH
Identification number 0123**Valid until 2026-12-14**
according to our valid EU Certificate
(No. G11 012974 0626)

Anlage I / Attachment I

Basic UDI-DI: 403923900000238936

Art.-Nr. / Art. No.	Produktname / Product name	Klasse / Class
4637100	Perifix® LOR	Is (sterile)
4637110	Perifix® LOR NRFit®	Is (sterile)
4638107	Perifix® LOR	Is (sterile)
4638110	Perifix® LOR NRFit®	Is (sterile)

Document amendment information

Version	Description of the changes
1.0	Initial Version under 2017/745 MDR

Title: Declaration of Conformity - 196-099-MDR - Perifix LOR Initiator: Sandra ? Staufenberg

This document is signed electronically in compliance with the B. Braun electronic signature policies and procedures by following persons:

UserName: Staufenberg, Sandra (stausade)
Title: Administrator Regulatory Affairs CoE Infusion & Pain Therapy
Date: Tuesday, 18 July 2023, 10:11 W. Europe Daylight Time
Meaning: Document signed as Author
=====

UserName: Seidel, Stefan (seidstde)
Title: Head of Regulatory Affairs CoE Infusion & Pain Therapy
Date: Tuesday, 18 July 2023, 13:26 W. Europe Daylight Time
Meaning: Approve Document
=====

UserName: Brand, Thomas (brantode)
Title: HC-QM-DE08 Vice President QM for non-active Medical Devices
Date: Thursday, 20 July 2023, 15:08 W. Europe Daylight Time
Meaning: Approve Document
=====