
APPLIMED SA

Zone industrielle
Route de Pra de Plan 1
CH-1618 CHATEL-ST-DENIS
Switzerland

TEL : +41 (0) 21 948 92 74
FAX : +41 (0) 21 948 90 36
E-mail : info@applimed.ch
Website : www.applimed.ch

PRODUCT DATA

I-Product Description & Classification			
I/1	Designation	Irrigation Cannula	
I/2	General Description	Irrigation Cannula, Straight, Luer lock, 80mm, 17G	
I/3	GMDN N°	30889	
I/4	Reference Number	8001738	
I/5	Class	☐ I ☐ Is ☒ IIa According to Directive 93/42/EEC - Annex IX, Rules 6	
I/6	Notify Body	**CE**0120	
I/7	Composition Raw Material	Stainless Steel 304/Cr	
I/8	Sterile	☒ Yes ☐ No If, yes ☒ ETO ☐ Gamma ☐ Steam	
I/9	First Packaging	Number of pieces in Packaging:	1 piece per peel pack
I/10	Picture		


II- Description of use and operating of device		
II/1	Intended use/indication:	This medical device is a surgical instrument used for the cleaning and irrigation of the wound. The olive protects the tissues against the needle tip. Single use and transient Use (<60min)
II/2	User:	Medical staff or professional
II/3	Contraindication of use	Not use, if the device is dirty or visual defects If packaging is damaged or open do not use. Sterile unless package is damaged.
II/4	Special requirement for use	n/a
II/5	Handling and storage conditions:	Standard conditions or atmospheric conditions

III- Supplementary information		
III/1	Compatibility with others medical devices (in the set)	n/a
III/2	Device contents animal substances	NO
III/3	Device contents medicinal substances	NO
III/4	Device comes into contact with substances	YES (usual substances used for treatment and cleaning of wounds)
III/5	Used with Accessories	NO
III/6	Used in combination with other Medical Device	NO

APPLIMED SA

Zone industrielle
Route de Pra de Plan 1
CH-1618 CHATEL-ST-DENIS
Switzerland

TEL : +41 (0) 21 948 92 74
FAX : +41 (0) 21 948 90 36
E-mail : info@applimed.ch
Website : www.applimed.ch



IV- MANUFACTURING

IV/1	Manufacturing Operations	☒ peel pack ☐ pouches ☒ boxes ☒ cartons ☒ cleaning ☒ sterilization	1 pieces per peel pack 100 peel pack per box 10 box per cartons
IV/2	Labelling: According to Directive 93/42/EEC Annex I § 13.3 and ISO 15223-1 • Name and description of product • Name and address of manufacturer • LOT N° • Article N° • Expiry date • Quantity • Sterile • Method of sterilisation • Indication for single use • Not use if the pack is opened • CE 0120	Example of a label **Aiguille boutonnée** 1 Aiguille boutonnée 80mm, 17G 100 Peelpacks/boîte 10 Boîtes/carton REF 8001738 2019-11 LOT 14.11.20 CE 0120 STERILE EO Applimed SA Pra de Plan 1 CH-1618 Châtel-St-Denis SAD (01)07640103631652(17)191120(10)141120	
IV/3	Place of sterilisation:	ETO gas sterilization according to ISO 11135-1, By CSSR (Centre de sterilisation de Suisse Romande)	