

Declaração CE de Conformidade

PMH - Produtos Médico Hospitalares, SA
Z.I. da Murteira LT9, Porto Alto,
2135-311 Samora Correia
Portugal

Declara que os seguintes dispositivos médicos:

TORNEIRAS E PROLONGAMENTOS PARA ADMINISTRAÇÃO DE SOLUÇÕES

Ref: AMT0031
AMT0032
AMT0033

Cumprem os requisitos essenciais do Anexo I da Diretiva 93/42/CEE do Conselho de 14 de junho de 1993 (alterada pela Diretiva 2007/47/CE, relativa aos dispositivos médicos, transposta para a lei Portuguesa e decretos aplicáveis.

A conformidade destes produtos de classe I estéril (regras de classificação do anexo IX da Diretiva acima mencionada) foi avaliada de acordo com o anexo V da Diretiva de acima sob a supervisão do Organismo Notificado Número 2797, BSI, Say Building, John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands.

EC Declaration of Conformity

PMH - Produtos Médico Hospitalares, SA
Z.I. da Murteira LT9, Porto Alto,
2135-311 Samora Correia
Portugal

Declare that following medical devices:

STOPCOCKS AND EXTENSION SETS

Ref: AMT0031
AMT0032
AMT0033

Comply with the essential requirements of the Annex I of the Directive 93/42/EEC of June 14, 1993 (amended by Directive 2007/47/EC concerning medical devices, transposed in Portuguese Law and the applicable de decrees.

The conformity of these products of class I sterile (classification rules of Annex IX of the above Directive) has been assessed according Annex V of the above Directive under the supervision of Notified Body Number 2797, BSI, Say Building, John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands.

Data/Date: 25/03/2021

Assinatura / Signature:

Sílvia Alexandra Fernandes
Diretor Qualidade / Quality Director
PMH - Produtos Médico Hospitalares, SA

P. M. H.
PRODUTOS MÉDICO HOSPITALARES S. A.
CONTRIBUINTE Nº 502 376 899
Z. I. MURTEIRA LT9
2135-311 SAMORA CORREIA
TEL.: + 351 263 650 680 FAX.: + 351 263 650 695

