Containerfilter für Martin-Container

Containerfilter aus Papier zum Einmalgebrauch für Martin Container nach ISO 11607-1. Der Containerfilter ist zum Einmalgebrauch und nicht wiederaufbereitbar.

Produktaufbau: Sterilisationspapier krepp weiß,
60 g/m² nach ISO 11607-1

Toxizität: eine toxische Wirkung ist nicht bekannt

Latex: latexfrei

Aufdruck: Ja
(**Druckerfarbe:** 1/0-fbg., schwarz, sterilisationsecht)

Indikator: mit Dampfindikator Typ 1, Farbumschlag von rosa nach dunkelbraun
entspricht der Norm: EN ISO 11140-1:2014 Sterilisation von Produkten für die Gesundheitsfürsorge – Chemische Indikatoren – Teil 1: Allgemeine Anforderungen

Sterilisationsverfahren:

STEAM (Dampf)	☒ geeignet	☐ ungeeignet
EO (Ethylenoxid)	☐ geeignet	☒ ungeeignet
FORM (Formaldehyd)	☐ geeignet	☒ ungeeignet
VH2O2 (Plasma)	☐ geeignet	☒ ungeeignet

Kompatibilitätshinweise:

AESCULAP:	☐ geeignet	☒ ungeeignet
WAGNER:	☐ geeignet	☒ ungeeignet
MARTIN:	☒ geeignet	☐ ungeeignet
MEDICON:	☐ geeignet	☒ ungeeignet

RDG-tauglich: Nein

Medizinprodukt: Klasse I

REF	Format in mm	Ecken	Lochstanzungen
01330	104 x 104	abgerundet	1 (lang)
01158	110 x 240	spitz	2 (rund)
01143	180 x 270	spitz	-
01321	180 x 255	spitz	-


Container filters for Martin-Containers

Disposable paper container filter for Martin containers according to ISO 11607-1. The container filter is disposable and non-recyclable.


Product structure:	sterilization crepe paper white, 60 g/m² according to ISO 11140-1		
Toxicity:	A toxic effect is not known		
Latex:	Not made with natural rubber latex		
Imprint:	Yes (**print colour:** 1/0-colour, black, sterilizable)		
Indicator:	with type 1 steam indicator, colour changes from pink to dark brown in accordance with: EN ISO 11140-1:2014 Sterilization of healthcare products - Chemical indicators - Part 1: General requirements		
Sterilization procedure:	STEAM (steam)	☒ suitable	☐ not suitable
	EO (Ethylene oxide)	☐ suitable	☒ not suitable
	FORM (Formaldehyde)	☐ suitable	☒ not suitable
	VH2O2 (Plasma)	☐ suitable	☒ not suitable
Compatibility notes:	AESCULAP:	☐ suitable	☒ not suitable
	WAGNER:	☐ suitable	☒ not suitable
	MARTIN:	☒ suitable	☐ not suitable
	MEDICON:	☐ suitable	☒ not suitable
RDG-suitable:	No		
Medical device:	Class I		

REF	Format in mm	Corners	Hole punching
01330	104 x 104	rounded	1 (long)
01158	110 x 240	pointed	2 (round)
01143	180 x 270	pointed	-
01321	180 x 255	pointed	-

01330



01158



01143/01321

